


Quotation Advert

Opening Date: 2019-07-31 
Closing Date: 2019-08-07 
Closing Time: 11:00

INSTITUTION DETAILS



Institution Name: Edendale hospital 
Province: KwaZulu-Natal
Department or Entity: Department of Health
Division or section: Central Supply Chain Management
Place where goods / services is required: Edendale Hospital
Date Submitted: 2019-07-30 

ITEM CATEGORY AND DETAILS

Quotation Number: ZNQ:
460/19-20
Item Category: Goods 
Item Description:
CATHETER FOLEY BALLON (2 WAY STANDARD) 12FG
Please note that no Emailed Quotations allowed

Quantity (if supplies) 1000 UNITS

COMPULSORY BRIEFING SESSION / SITE VISIT

Select Type: Not Applicable 
Date : 
Time:
Venue: Edendale Hospital Supply chain managent

QUOTES CAN BE COLLECTED FROM: Edendale Hospital Supply Chain Management

QUOTES SHOULD BE DELIVERED TO: Edendale Hospital blue tender box at the main gate

ENQUIRIES REGARDING THE ADVERT MAY BE DIRECTED TO:

Name: Mrs Busie Mkhize
Email: busisiwe.mkhize4@kznhealth.gov.za
Contact Number: 033 395 4250
Finance Manager Name: Mr D. Thangalan

Finance Manager Signature: _____


No late quotes will be considered



health

Department:
Health
PROVINCE OF KWAZULU-NATAL

DIRECTORATE: EDENDALE HOSPITAL

Physical Address: Main Road Edendale, Edendale
Postal Address: Private Bag X509, Plessislaer, 3216
Tel: 033 3954251 Fax: 033 3954270 Email: anil.sookdavu@kznhealth.gov.za
www.kznhealth.gov.za

SUPPLY CHAIN MANAGEMENT

**SPECIFICATION CATHETER, URETHRAL, FOLEY, TWO-WAY SILICONE ELASTOMER,
COATED LATEX 16FG**

SECTION A – General

Clause	Clause Description	Bidders Remarks
G1	Bidders must quote on the supply and delivery of the item.	
G2	SABS/SANS 51616 proof of certification or any other relevant certification documentation must be provided with the quotation. G2.1 Failure to provide proof of certification may render your bid as not valid.	
G3	Bidders must provide a sample of the item quoted for.	
G3.1	Please state brand name of item quoted for:	
G4	Preference will be given to a brand that all clinicians have been trained and skilled to use. Any other brands will have to be trialed and the end-user training given at no extra cost to the final bid price.	
G5	Bidder to provide an operating manual which gives clear instructions.	
G8	Bidder to ensure that the following information is clearly indicated on each of the units. G8.1 The correct description of the contents. G8.2 The correct size of the contents. G8.3 The Lot/batch number. G8.4 The manner in which sterilization was done. G8.5 Name and contact details of the manufacturer. G8.6 The expiry date. G8.7 Bidder to note that the expiry date must not be less than 18 months. Failing this the delivery will not be accepted and will have to be returned at the expense of the supplier. G8.8 Item code.	
G9	Bidder to provide references of other health institutions or clinicians currently utilizing the product offered.	
G10	Failure to return this specification document fully completed and signed may render your bid as not valid and will not be considered.	

SECTION B – Technical Specification

Clause	Clause Description	Bidders Remarks
T1	1.1 The required urethral catheter must be a two way, silicone elastomer-coated latex.	
T2	Proximal end: One tapered end and one balloon port with a non return valve and fitting for a syringe.	
T3	Distal end: closed round tip with two lateral eyes distal to a 5ml capacity balloon.	
T4	Size: 16FG (Outer Diameter 5.3mm)	
T5	Must be packed individually in a peel packed wrapper, double wrapped – inner pouch unsealed.	
T6	Must be single use.	
T7	Must be sterile	

SECTION C – Confirmation

Compulsory to be completed by the bidder and returned with quotation:

I,, from do hereby	
(Print Name)	(Name of Business)
acknowledge that I have read and understand the specifications as laid out above and will ensure that the quotation price submitted will address all the requirements as stipulated.	
..... / /
Signature	Date